

## Proposal for using the polarimeter data for the spring 2004 conferences

- TPOL has large scale uncertainty (5 – 10%) due to changes of detector, beam parameters (emittance). Offline analysis not yet fully understood
- LPOL detector not changed, but systematic error not yet evaluated
- Proposal: use LPOL numbers for spring conferences. Use  $\Delta P = 5\%$  as preliminary systematic uncertainty