### Software Alignment for Tracking Detectors

January 17<sup>th</sup> 2005

Volker Blobel – Universität Hamburg

#### Abstract

Tracking detectors in high energy physics experiments require an accurate determination of a large number of alignment parameters in order to allow a precise reconstruction of tracks and vertices. In addition to the initial optical survey and corrections for electronics and mechanical effects the use of tracks in a special software alignment is essential. Several different methods are in use, ranging from simple residual-based procedures to complex fitting systems with many thousands of parameters. The methods are reviewed with respect to their mathematical basis and accuracy, and to aspects of the practical realization.

- 1. Alignment and calibration
- 2. Alignment of a toy detector
- 3. Linear least squares and matrix equations
- 4. The MILLEPEDE algorithm

- 5. Drift chamber calibration
- 6. Vertex detector alignment
- 7. Conclusions and summary

Keys during display: enter = next page;  $\rightarrow$  = next page;  $\leftarrow$  = previous page; home = first page; end = last page (index, clickable); C- $\leftarrow$  = back; C-N = goto page; C-L = full screen (or back); C-+ = zoom in; C-0 = fit in window; C-M = zoom to; C-F = find; C-P = print; C-Q = exit.

# 1. Alignment and calibration

What is alignment and calibration? ... a web search

#### Alignment:

**1.** an adjustment to a line; arrangement in a straight line.

**2.** the line or lines so formed.

3. the proper adjustment of the components of an electronic circuit, machine, etc., for coordinated functioning: The front wheels of the car are out of alignment.
4. a state of agreement or cooperation among persons, groups, nations, etc., with a common cause or viewpoint.

5. a ground plan of a railroad or highway.

**6.** (Archaeol.) a line or an arrangement of parallel or converging lines of upright stones or menhirs.



Alignement at Kermario (Bretagne)

**Purpose of instrument calibration:** Instrument calibration is intended to *eliminate or reduce bias* in an instrument's readings over a range for all continuous values. For this purpose, *reference standards* with known values for selected points covering the range of interest are measured with the instrument in question. Then a *functional relationship* is established between the values of the standards and the corresponding measurements [...from NIST].

## Alignment/calibration

Calibration seems to be the more general term, while alignment is related to geometrical calibration.

Methods in HEP: many papers and collaboration reports on the Web, with a lot of HEP folklore (residuals, pulls, fit errors, inversion of a rectangular matrix, ... with the usual gradient method, MINUIT, typically  $\mathcal{O}(10\mu m)$ , chi-square minimization, constraints, convergence recognized by small change/iteration, < 20 iterations, ... our method is successful).

Sometimes it is difficult to understand what is really done: "The detector parameters are found via a  $\chi^2$  minimization of the residuals."

Aim of **alignment/calibration** with tracks, after an initial optical survey and corrections for electronics and mechanical effects:

- eliminate or reduce bias in detector data;
- increase precision of reconstructed tracks and vertices;
- improve track and vertex recognition, reduce  $\chi^2$  of the fits;
- essential for accurate vertex detectors with potential precision of a few  $\mu$ m for e.g. heavy quark physics.

Alignment/calibration requires to *understand* the detector (functional relationship) and to optimize thousands of parameters – not a simple task. **Residuals** play a central role.

Test of alignment method with a MC toy track detector model:

- 10 planes of tracking chambers, 1 m high, 10 cm distance, no magnetic field;
- accuracy  $\sigma \approx 200 \mu \text{m}$ , with efficiency  $\epsilon = 90\%$ ;
- plane 7 sick: accuracy σ ≈ 400µm, with efficiency ε = 10%;
- 10 000 tracks with 82 000 hits available for alignment;
- Misalignment: the vertical position of the chambers are displaced by  $\approx 0.1$ cm (normal distributed).



The first alignment attempt is based on the distribution of hit residuals:

- A straight line is fitted to the track data.
- The residuals (= measured vertical coordinate minus fitted coordinate) are histogrammed, separately for each plane.



• The mean value of the residuals is taken as correction to the vertical plane position.

This is the standard method used in many experiments.

## Alignment using iterative track fitting



Properties of the method:

- simple, only basic operations  $+ \times /$  used, no matrices, *n*-tuples can be used;
- "simple, intuitive, quick", "method very robust" (i.e. small changes expected from residuals);
- but based on biased straight line fit!
- Is the method convergent, if applied iteratively, and if convergent, what is the order (linear of quadratic) of convergence?

ID	true shift	determined	mean residual
1	0.1391	0.0727	$0 \pm 150$
2	0.1345	0.0786	$0 \pm 189$
3	0.0000	-0.0453	$0 \pm 234$
4	-0.0756	-0.1102	$0 \pm 244$
5	-0.1177	-0.1422	$0 \pm 205$
6	0.0610	0.0475	$0 \pm 150$
$\overline{7}$	0.0130	0.0114	$0 \pm 464$
8	0.0886	0.0968	$0 \pm 255$
9	0.0000	0.0186	$0 \pm 149$
10	-0.0467	-0.0176	$0 \pm 143$

After 30 iterations ...



red circle = true shift (displacement)

blue disc = displacement, determined from residuum

Large changes in first iteration, small changes in second iteration, almost no progress afterwards.

The result is not (yet) encouraging!

The reason for non-convergence is simple:

Two degrees of freedom are undefined: a simultaneous shift and a rotation of all planes! (This simple fact is not always mentioned in reports on the method!)

Improvement for second residual attempt:

Fix the displacement (i.e. displacement = 0) of two planes, which are assumed to be carefully aligned externally (e.g. planes 3 and 9).

Other possibilities are:

- Use only fixed planes (planes 3 and 9) in the fit, and determine the residuals of other planes;
- for the determination of the displacement of a certain plane use all other planes in the fit.

These possibilities are in fact used by several collaborations!

ID	true shift	determined	mean residual
1	0.1391	0.1391	$-1 \pm 150$
2	0.1345	0.1344	$0 \pm 189$
3	0.0000	0	$2 \pm 234$
4	-0.0756	-0.0758	$0 \pm 244$
5	-0.1177	-0.1183	$0 \pm 205$
6	0.0610	0.0607	$0 \pm 150$
7	0.0130	0.0140	$0 \pm 464$
8	0.0886	0.0888	$0 \pm 255$
9	0.0000	0	$0 \pm 149$
10	-0.0467	-0.0469	$0 \pm 143$

After 30 iterations with planes 3 and 9 fixed (displacement = 0) ...

0.2

-0.2

$$\label{eq:circle} \begin{split} \mathrm{red}\ \mathrm{circle} &= \mathrm{true}\ \mathrm{shift}\ (\mathrm{displacement})\\ \mathrm{blue}\ \mathrm{disc} &= \mathrm{displacement},\ \mathrm{determined}\ \mathrm{from}\ \mathrm{residuum} \end{split}$$



10

Shifts from residuals - iteration 30

5

Residual-based methods work with biased results. Can the bias be avoided by an improved fit?

**Yes**: include the alignment parameters in the parameters fitted in track fits – requires a simultaneous fits of many tracks, with determination of (global) alignment parameters and (local) track parameters.

... a linear least squares problem of  $m = 82\ 000$  equations (measurements) and  $n = 20\ 010$  parameters with  $n \ll m$ , which requires the solution of a matrix equation with 20010-by-20010 matrix.

 $\ldots$  a nice problem!

ID	true shift	determined	ρ	mean residual
1	0.1391	$0.1393\pm0.004$	0.68	$0 \pm 150$
2	0.1345	$0.1346\pm0.003$	0.66	$0 \pm 189$
3	0.0000			$0 \pm 234$
4	-0.0756	$-0.0756 \pm 0.003$	0.58	$0 \pm 244$
5	-0.1177	$-0.1182 \pm 0.003$	0.53	$0 \pm 205$
6	0.0610	$0.0608 \pm 0.003$	0.50	$0 \pm 150$
7	0.0130	$0.0141\pm0.007$	0.20	$0 \pm 464$
8	0.0886	$0.0888 \pm 0.003$	0.53	$0 \pm 255$
9	0.0000			$0 \pm 149$
10	-0.0467	$-0.0469 \pm 0.003$	0.57	$0 \pm 143$

After one step (with planes 3 and 9 fixed at displacement = 0) ...

 $(\rho = \text{global correlation coefficient})$ 

red circle = true shift (displacement)blue disc = displacement, determined in fit



One step is sufficient: 1. step  $\Delta \chi^2 = 1.277 \times 10^6$  2. step  $\Delta \chi^2 = 1.159 \times 10^{-5}$ 

But how can this problem be solved in less than a second?

Improvement: include, in addition, corrections to the drift velocities for each plane:  $\Delta v_{\text{drift}}/v_{\text{drift}}$ 

$$y_i \cong a_1^{\text{local}} + a_2^{\text{local}} \cdot x_i + a_j^{\text{global}} + \ell_{\text{drift},i} \cdot \left(\frac{\Delta v_{\text{drift}}}{v_{\text{drift}}}\right)_j \qquad a_j^{\text{global}} = \text{shift for plane}$$
  
 $\left(\frac{\Delta v_{\text{drift}}}{v_{\text{drift}}}\right)_j = \text{relative } v_{\text{drift}} \text{ difference}$ 

#### reduction of residual $\sigma$ by 30 - 40 %

ID	true shift	determined	ρ	$\Delta v_{\rm drift}/v_{\rm drift}$	determined	ρ	mean residual
1	0.1391	$0.1393 \pm 0.004$	0.68	0.0020	$0.0019 \pm 0.0002$	0.016	$0 \pm 119$
2	0.1345	$0.1346\pm0.003$	0.66	-0.0153	$-0.0150 \pm 0.0002$	0.020	$0 \pm 128$
3	0.0000			0.0193	$0.0194 \pm 0.0002$	0.017	$0 \pm 137$
4	-0.0756	$-0.0756 \pm 0.003$	0.58	0.0200	$0.0197 \pm 0.0002$	0.013	$0 \pm 139$
5	-0.1177	$-0.1182 \pm 0.003$	0.53	-0.0138	$-0.0136 \pm 0.0002$	0.013	$0 \pm 141$
6	0.0610	$0.0608 \pm 0.003$	0.50	0.0003	$0.0004 \pm 0.0002$	0.019	$0 \pm 139$
7	0.0130	$0.0141\pm0.007$	0.20	-0.0306	$-0.0303 \pm 0.0006$	0.038	$0 \pm 348$
8	0.0886	$0.0888 \pm 0.003$	0.53	0.0237	$0.0238 \pm 0.0002$	0.018	$0 \pm 134$
9	0.0000			-0.0044	$-0.0044 \pm 0.0002$	0.008	$0 \pm 127$
10	-0.0467	$-0.0469 \pm 0.003$	0.57	0.0021	$0.0019 \pm 0.0002$	0.013	$0 \pm 117$

... this would be rather difficult with a pure residual-based method.

The next improvement would be the introduction of wire  $T_0$ 's – additional  $10 \times 25 = 250$  parameters.

## 3. Linear least squares and matrix equations

## Gauss, Laplace and Lagrange



Johann Carl Friedrich Gauss (1777 - 1855)

 $\begin{array}{l} \text{Pierre-Simon Laplace} \\ (1749-1827) \end{array}$ 

Joseph-Louis Lagrange (1736 - 1813)

V. Blobel – University of Hamburg

Alignment for tracking detectors

**Measurement:** data  $y_i$ , i = 1, 2, ..., m, measured independently and without bias at coordinate  $x_i$ , with standard deviation  $\sigma_i$ .

Mathematical model: data y are described by the mathematical model

$$y = f(x; \boldsymbol{a})$$

with a function f(x; a), depending on the coordinate x and parameters  $a_j$ , j = 1, 2, ..., n.

The function  $f(x; \mathbf{a})$  is linear or can be linearized. The method of linear least squares can be used to perform one step in improving the parameter estimate  $\mathbf{a}_{\ell}$  by a correction vector  $\Delta \mathbf{a}$ :

$$y_i \cong f(x_i; \boldsymbol{a}_\ell) + \sum_{j=1}^n d_j \Delta a_j \quad \text{with} \quad d_j = \left. \frac{\partial f(x)}{\partial a_j} \right|_{x=x_i}$$

The principle of least squares requires to minimize the residuals

$$r_i = y_i - f(x_i; \boldsymbol{a}_\ell) \cong \boldsymbol{d}_i^{\mathrm{T}} \Delta \boldsymbol{a}$$

taking into account the measurement accuracy  $\sigma_i$ .

#### The standard linear least squares method II

The least squares requirement minimize 
$$S = \sum_{i=1}^{m} w_i r_i^2$$
 with weight  $w_i = \frac{1}{\sigma_i^2}$  requires to solve the system of  $m$  linear equations (normal equations) or matrix equation

$$C \Delta a = b$$

with the symmetric *n*-by-*n* matrix  $\boldsymbol{C}$  and the *n*-vector  $\boldsymbol{b}$ :

$$\boldsymbol{C} = \sum_{i=1}^{m} w_i \, \boldsymbol{d}_i \, \boldsymbol{d}_i^{\mathrm{T}} \qquad \boldsymbol{b} = \sum_{i=1}^{m} w_i \, r_i \, \boldsymbol{d}_i$$

This completes one step of the iteration

$$\boldsymbol{a}_{\ell+1} := \boldsymbol{a}_{\ell} + \Delta \boldsymbol{a} = \boldsymbol{a}_{\ell} + \boldsymbol{C}^{-1} \boldsymbol{b}$$
 and  $\ell := \ell + 1$ 

No iteration is necessary for a function, which is linear in the parameters. After convergence  $a \equiv a_{\ell}$  is an *unbiased estimate* of the parameter vector, with covariance matrix given by the inverse matrix  $C^{-1}$  (error propagation).

residual 
$$r_i = y_i - f(x_i; \boldsymbol{a})$$
 pull  $p_i = \frac{y_i - f(x_i; \boldsymbol{a})}{\sqrt{\sigma_i^2 - \sigma_{i, \text{fit}}^2}}$  with  $\sigma_{i, \text{fit}}^2 = \boldsymbol{d}_i^{\mathrm{T}} \boldsymbol{C}^{-1} \boldsymbol{d}_i$ 

Qualitative investigations using histograms of the residuals and quantitative tests by checking pull distributions:  $p_i \sim N(0, 1)$  expected.

## 4. The MILLEPEDE algorithm

Simultaneous least squares fit of all global and all local parameters (i.e. all tracks).

k'th track: 
$$y_i \cong f(x_i; \boldsymbol{a}^{\text{global}}, \boldsymbol{a}^{\text{local}}) + \left(\boldsymbol{d}_i^{\text{global}}\right)^{\mathrm{T}} \Delta \boldsymbol{a}^{\text{global}} + \left(\boldsymbol{d}_i^{\text{local}}\right)^{\mathrm{T}} \Delta \boldsymbol{a}_k^{\text{local}}$$

The complete matrix equation for global and local parameters includes sums over track index k and contains many matrices: *n*-by-*n* matrices C for *n* global parameters and *m*-by-*m* matrices  $C_k^{\text{local}}$  and *n*-by-*m* matrices  $H_k^{\text{global-local}}$ 



If the  $\boldsymbol{H}_{k}^{\text{global-local}}$  are neglected, the complete equation decays into 1+K independent matrix equations. But the solution  $\Delta \boldsymbol{a}^{\text{global}}$  can be calculated without approximation with a great simplification:  $\Rightarrow$  A square *n*-by-*n* matrix C and its inverse,  $C^{-1} \equiv B$  can be partitioned into submatrices

Submatrices  $C_{11}$  and  $B_{11}$  are *p*-by-*p* square matrices and submatrices  $C_{22}$  and  $B_{22}$  are *q*-by-*q* square matrices, with p + q = n. A complete set of equations for the submatrices of the inverse matrix B can be derived from the matrix product  $C B = \mathbf{1}_n$ :

$m{C}_{11} \ m{B}_{11} + m{C}_{12} \ m{B}_{21} = m{1}_p$	$m{B}_{11} = (m{C}_{11} - m{C}_{12}m{C}_{22}^{-1}m{C}_{21})^{-1}$
$m{C}_{11} \; m{B}_{12} + m{C}_{12} \; m{B}_{22} = m{0}$	$m{B}_{12} = -m{B}_{11}m{C}_{12}m{C}_{22}^{-1}$
$m{C}_{21} \; m{B}_{11} + m{C}_{22} \; m{B}_{21} = m{0}$	$m{B}_{21} = -m{C}_{22}^{-1}m{C}_{21}m{B}_{11}$
$m{C}_{21} \; m{B}_{12} + m{C}_{22} \; m{B}_{22} = m{1}_q$	$\boldsymbol{B}_{22} = \boldsymbol{C}_{22}^{-1} + \boldsymbol{C}_{22}^{-1} \boldsymbol{C}_{21} \boldsymbol{B}_{11} \boldsymbol{C}_{12} \boldsymbol{C}_{22}^{-1}$

Only the inversion of a *p*-by-*p* and of a *q*-by-*q* matrix  $(C_{22})$  is required.

The special structure of a matrix to be inverted may allow a significant reduction of the computational effort of the inversion by the method of partitioning (e.g.  $C_{22}$  diagonal). Further simplification for symmetric matrices.

Partitioning of the whole matrix equation including the vectors  $\Delta \boldsymbol{a}$  and  $\boldsymbol{b}$ :

$$\begin{pmatrix} \mathbf{C}_{11} & \mathbf{C}_{12} \\ \hline \mathbf{C}_{21} & \mathbf{C}_{22} \end{pmatrix} \begin{pmatrix} \Delta \mathbf{a}_1 \\ \hline \Delta \mathbf{a}_2 \end{pmatrix} = \begin{pmatrix} \mathbf{b}_1 \\ \hline \mathbf{b}_2 \end{pmatrix} \qquad \qquad \begin{pmatrix} \Delta \mathbf{a}_1 \\ \hline \Delta \mathbf{a}_2 \end{pmatrix} = \begin{pmatrix} \mathbf{B}_{11} & \mathbf{B}_{12} \\ \hline \mathbf{B}_{21} & \mathbf{B}_{22} \end{pmatrix} \begin{pmatrix} \mathbf{b}_1 \\ \hline \mathbf{b}_2 \end{pmatrix}$$

Special case  $C_{12} = 0$ :

independent solutions  $\Delta \widetilde{\boldsymbol{a}}_1 = \left( \boldsymbol{C}_{11}^{-1} \right) \boldsymbol{b}_1 \qquad \Delta \widetilde{\boldsymbol{a}}_2 = \left( \boldsymbol{C}_{22}^{-1} \right) \boldsymbol{b}_2 \; .$ 

General case  $C_{12} \neq 0$ : solution  $\Delta a_1$  can be expressed by  $\Delta \tilde{a}_2$ 

$$\Delta \boldsymbol{a}_{1} = \boldsymbol{B}_{11}\boldsymbol{b}_{1} + \boldsymbol{B}_{12}\boldsymbol{b}_{2} = \boldsymbol{B}_{11}\boldsymbol{b}_{1} + \boldsymbol{B}_{11}\boldsymbol{C}_{12}\left(\boldsymbol{C}_{22}^{-1}\boldsymbol{b}_{2}\right) = \boldsymbol{B}_{11}\left(\boldsymbol{b}_{1} - \boldsymbol{C}_{12}\Delta\widetilde{\boldsymbol{a}}_{2}\right)$$
$$= \left(\boldsymbol{C}_{11} - \boldsymbol{C}_{12}\boldsymbol{C}_{22}^{-1}\boldsymbol{C}_{12}^{T}\right)^{-1}\left(\boldsymbol{b}_{1} - \boldsymbol{C}_{12}\Delta\widetilde{\boldsymbol{a}}_{2}\right)$$

i.e. the full solution  $\Delta a_1$  can be obtained by "corrections" to  $C_{11}$  and  $b_1$  (without calculating  $\Delta a_2$ ).

This equation can be applied repeatedly with a problem – opens the possibility to solve problems with a huge number of parameters, if the interest is in a sub-vector of the parameters

V. Blobel – University of Hamburg

Alignment for tracking detectors

For each track in a loop, on all tracks:

- 1. Track- or other fit: perform fit by finding the best local parameter values for the actual track until convergence with determination of the covariance matrix V of the local parameters
- 2. Derivatives: calculate for all hits (index *i*) the vectors of derivatives  $d_i^{\text{local}}$  and  $d_i^{\text{global}}$  for all local and the relevant global parameters, and update matrices:

$$oldsymbol{C} := oldsymbol{C} + \sum_{i} w_i oldsymbol{d}_i^{ ext{global}} \left(oldsymbol{d}_i^{ ext{global}}
ight)^{ ext{T}}$$
  $oldsymbol{b} := oldsymbol{b} + \sum_{i} w_i r_i oldsymbol{d}_i^{ ext{global}}$   $oldsymbol{H} = \sum_{i} w_i oldsymbol{d}_i^{ ext{global}} \left(oldsymbol{d}_i^{ ext{local}}
ight)^{ ext{T}}$   
and finally for the track  $oldsymbol{C} := oldsymbol{C} - oldsymbol{H} oldsymbol{H}^T$   
The two 'blue' equations transfer the 'local' information to the global parameters.

After the loop on all tracks the complete information is collected; now the matrix equation for the global parameters has to be solved:

$$\Delta \boldsymbol{a}^{ ext{global}} = \boldsymbol{C}^{-1} \boldsymbol{b}$$

Note: matrices C and vectors b from several processors can be simply added to get combined result.

Undefined degrees of freedom can be avoided by adding parameter constraint equations of type

 $g(\boldsymbol{a}) = 0$ 

e.g. "zero average displacement", or "zero rotation of the whole detector".

There are two possibilities:

- fix certain parameters simply set rows/columns of the fixed parameter to zero – the matrix inversion program will take care of that.
- add equality constraint equation append linearized Lagrange multiplier equation  $\lambda \left( g(\boldsymbol{a}) + \boldsymbol{g}^{\mathrm{T}} \cdot \Delta \boldsymbol{a} = 0 \right)$  with  $\boldsymbol{g} = \partial g(\boldsymbol{a}) / \partial \boldsymbol{a}$ :



V. Blobel: Experience with Online Calibration Methods, Contribution to CHEP'97, Berlin 1997 (including the Millepede method), not accepted.

Used/tested in H1(1997), CDF(2001), Hera-B(?), ZEUS(?), Atlas(?)  $\dots$ 

O. Behnke: Directions in Tracking (talk), H1 Collaboration meeting at Zürich, 1999

V. Blobel, Linear Least Squares Fits with a Large Number of Parameters, (2000), and 1. Update (2000) http://www.desy.de/~blobel including Fortran code. C++ interface for Millepede by Jesko Merkel: http://www.desy.de/~jmerkel/cmillepede.html

V. Blobel and C. Kleinwort: A New Method for the High-Precision Alignment of Track Detectors, Proc. of the Conference on: Advanced Statistical Techniques in Particle Physics, Durham, March 18th - 22nd, 2002. arXiv-hep-ex/0208021

The algorithm is general and can be applied to other problems with large number of (global) parameters and a huge number of measurements with local parameters.

Principle of reducing matrix size (perhaps) used earlier:

Schreiber, O. (1877): Rechnungsvorschriften für die trigonometrische Abteilung der Landesaufnahme, Ausgleichung und Berechnung der Triangulation zweiter Ordnung. Handwritten notes. Mentioned in W. Jordan (1910): Handbuch der Vermessungskunde, Sechste erw. Auflage, Band I, Paragraph III: 429-433. J.B.Metzler, Stuttgart.

### The MILLEPEDE program

The matrix-equations require about 100 lines of code (including optimization for sparse matrices); in addition the routine for the solution of the matrix equation has also about 100 statements [subroutine MMPEDE – Mini MILLEPEDE].

The more complete program MILLEPEDE on the Web has  $\approx 1700$  lines of code and includes iterations. Outliers are present and have to be removed by cuts – but initial cuts have to be wide in order to accept badly misaligned hits.

Unfortunately this requires iterations e.g. in MILLE-PEDE with

- $10\sigma$  cut in the first iteration,
- decreasing cut value in further iterations, and
- $3\sigma$  cut in the last iteration.



Z-residuals (units are 10  $\mu$ m) versus  $\phi$  in degrees for a LEP vertex detector – before and after MILLEPEDE alignment.

The region around  $\phi = 180^{\circ}$  with initially very large deviations (deviation  $\approx 10\sigma$ ) is moved to small values.

M. J. Fero et al., Performance of the SLD central drift chamber, Nucl. Instr. methods A 367 (1995) 111-114

Drift velocity calibration  $v_{\text{drift}}$ : first estimate of relation between  $v_{\text{drift}}$  and  $\boldsymbol{E}$ -field based on electrostatic model; relation iteratively corrected by minimizing the fit residuals as a function of drift distance.

Note:  $v_{\rm drift} \propto 1/{\rm pressure}$ . Tracks are used to measure changes in  $v_{\rm drift}$ , which is allowed to be one of the variables in the track fit, and averaging of values from individual tracks.

Claus Kleinwort et al., Detailed calibration of H1 drift chamber using MILLEPEDE with about 1400 parameters.

Examples:

- *common* alignment of the drift chamber and the silicon detector;
- for both CJC1 and CJC2 14 global parameters representing an overall shift or tilt are introduced;
- local variations of the drift velocity  $v_{\text{drift}}$  for cells halfs and layers halfs are observed, which are parametrized by 180 + 112 corrections, which change with the HV configuration;
- for each wire group (8 wires) corrections to  $T_0$  are introduced (330 corrections).

row.	number	parameter	σ	unit
1	2	$\Delta x$	1	$\mu \mathrm{m}$
2	2	$\Delta x / \Delta z_r$	2	$\mu m$
3	2	$\Delta y$	1	$\mu m$
4	2	$\Delta y/\Delta z_r$	2	$\mu m$
5	2	$\Delta \varphi$	10	$\mu$ rad
6	2	$\Delta \varphi / \Delta z_r$	10	$\mu$ rad
7	2	$\Delta \alpha_{ m Lor}$	100	$\mu$ rad
8	2	$\Delta v_{\rm drift-}/v_{\rm drift}$	$10^{-5}$	
9	2	$\Delta v_{\rm drift+}/v_{\rm drift}$	$10^{-5}$	
10	2	$\Delta T_0  imes v_{ m drift}$	< 1	$\mu { m m}$
11	2	wire staggering in wire plane	few	$\mu \mathrm{m}$
12	2	wire staggering perp wire plane	few	$\mu m$
13	2	sagging in wire plane	few	$\mu { m m}$
14	2	sagging perp. wire plane	few	$\mu { m m}$
15	180	$\Delta v_{\rm drift}/v_{\rm drift}$ per cell half	few	$10^{-4}$
16	112	$\Delta v_{\rm drift}/v_{\rm drift}$ per layer half	few	$10^{-4}$
17	330	$\Delta T_0 \times v_{\text{drift}}$ per group	10	$\mu { m m}$
18	56	wire position in driftdir. per layer	10	$\mu { m m}$
19	56	$\Delta T_0 \times v_{\rm drift}$ per layer	10	$\mu { m m}$
20	56	wire pos. perp. driftdir. per layer	few 10	$\mu { m m}$
21	112	$\Delta v_{\rm drift}/v_{\rm drift}$ for $I_e/50~{\rm mA}$	few $10^{-4}$	
22	90	$\Delta v_{\rm drift}/v_{\rm drift}$ per layer	few $10^{-4}$	
23	90	$\Delta y_W$ per layer	few 10	$\mu { m m}$
24	90	$\Delta y_W$ per layer <sup>2</sup>	few $10$	$\mu { m m}$
25	64	$\Delta$ in ladder	few	$\mu \mathrm{m}$
26	64	$\Delta$ perp. ladder	few	$\mu \mathrm{m}$
27	64	rel. $\Delta$ in ladder $(z_r)$	few	$\mu \mathrm{m}$
28	64	rel. $\Delta$ perp. ladder $(z_r)$	10	$\mu { m m}$
29	64	rel. $\Delta$ perp. ladder $(\varphi)$	few	$\mu { m m}$

Alignment parameters determined in the **Millepede** fit. The **Millepede** accuracy is given by a standard deviation  $\sigma$ .

#### H1 Drift chamber alignment and calibration



## 6. Vertex detector alignment

Planar sensors (silicon pixel or strip detectors): local (sensor) coordinates q = (u, v, w) and global detector coordinates r = (x, y, z) are transformed by

$$\boldsymbol{q} = \boldsymbol{R} \left( \boldsymbol{r} - \boldsymbol{r}_0 \right)$$
  $\boldsymbol{R} = \text{ nominal rotation}, \quad \boldsymbol{r}_0 = \text{ nominal position}$ 

After alignment the transformation becomes (with  $\Delta q = (\Delta u, \Delta v, \Delta w)$ )

$$\boldsymbol{q}^{c}=\boldsymbol{R}_{lpha}\boldsymbol{R}_{eta}\boldsymbol{R}_{\gamma}\;\boldsymbol{R}\left(\boldsymbol{r}-\boldsymbol{r}_{0}
ight)-\Delta \boldsymbol{q}$$

Six parameters are required for each individual detector element, out of which three parameters (two translations, one rotation) are very sensitive.

Detector	nr of elements	nr of parameters
SLD	96	$96 \times 6 = 576$
Aleph	144	$144 \times 6 = 864$
Delphi	24	72
CDF	352	$352 \times 3 = 1056$
Atlas		
CMS	$\approx 20000, \ge 13000$	$\approx 100000$

Atlas plans/ideas: system of linear equations with  $30000 \times 30000$  matrix; with hardware: 16 processors (64 bits achitecture) with 900 Mb memory for each processor, ScaLAPACK.

#### H1 calibration

805.4

-0.3628

62.42

800 1000 Δd<sub>CA</sub> [µm]

#### ... from O. Behnke, Directions in Tracking (Talk), Collaboration Meeting in Zürich 1999



### Internal alignment of the SLD vertex detector

D. J. Jackson, D. Su and F. J. Wickens, Internal alignment of the SLD vertex detector using a matrix singular value decomposition technique. Nuclear Instr. Methods A 491 (2002) 351-365

Classification of types of tracking constraints  $\rightarrow$ 

Good quality tracks were constrained to pass through two of the CCD hits with the corresponding residual measued to the third, reference, CCD.



For each type of residuals, *n*-tuples were accumulated containing the deviation. Data in the *n*-tuples were fitted to the functional form given below to determine the coefficients and their covariance matrix (using MINUIT).

CCD shape corrections were taken into account from the optical survey data of the CCD surfaces  $(2108 \rightarrow 5026 \text{ coefficients})$ .

Туре	Functional form		NI	Nc
Shingles	$\delta_z = s_1^{\parallel} + s_2^{\parallel} \tan \lambda + s_3^{\parallel} \tan^2 \lambda$		96	288
	$\delta_{L\phi} = s_1^{\perp} + s_2^{\perp} \tan \lambda$		96	192
Doublets	$\delta_z = d_1^{  } + d_2^{  } L_\phi$		48	96
	$\delta_{L\phi} = d_1^\perp + d_2^\perp L_\phi + d_3^\perp L_\phi^2$		48	144
Triplets	$\delta_{z} = t_{1}^{\parallel} + t_{2}^{\parallel} \tan \lambda + t_{3}^{\parallel} \tan^{2} \lambda + t_{4}^{\parallel} L_{\phi} \tan \lambda + t_{5}^{\parallel} L_{\phi}$		80	400
	$\delta_{L\phi} = t_1^{\perp} + t_2^{\perp} L_{\phi} + t_3^{\perp} L_{\phi}^2 + t_4^{\perp} L_{\phi} \tan \lambda + t_5^{\perp} \tan \lambda$		80	400
Pairs	$\delta_{rz} = p_1^{\parallel} + p_2^{\parallel} \tan \lambda + p_3^{\parallel} \tan^2 \lambda$		28	84
	$\delta_{r\phi} = p_1^{\perp} + p_2^{\perp} \tan \lambda$		28	56
	$\delta_{\phi} = p_1^{\phi} + p_2^{\phi} \tan \lambda$		28	56
CDC angle	$\delta_{\lambda} = c_1^{\lambda} + c_2^{\lambda} \tan \lambda + c_3^{\lambda} \tan^2 \lambda$		56	168
	$\delta_{\phi} = c_1^{\phi} + c_2^{\phi} \tan \lambda$		56	112
IP constraint	$\delta_{r\phi}=i_{1}^{\perp}+i_{2}^{\perp} an\lambda$		56	112
		Total	700	2108

Functional forms for fitting the various residual types. The superscripts || and || indicate coefficients of fits to residuals measured

Functional forms for fitting the various residual types.



Examples for residual fits, here as a function of  $\tan \lambda$  in a layer before the alignment.

In total there are 2108 coefficients from 700 residual fits (exercise in book-keeping).

Table 1

Taking into account the covariance matrices of the residual fits, 866 alignment corrections were determined from 5026 coefficients from the residual fits, using SVD techniques for the least-squares fit minimization.

Note: nr of coefficients determined by parametrization types, not by nr of events.

Only a single iteration is required.

With the aligned geometry the design performance is achieved  $\rightarrow$ 



Triplet residuals obtained with the original survey geometry and after alignment.

From the paper published in 2002:

"... for the VXD3 alignment (inversion of sparse matrices of order  $5000 \times 1000$  elements) using double precision arithmetic in modest times on a standard workstation. Although the size of matrices involved appears daunting only  $\approx 1\%$  or  $\approx 35,000$  elements of the final  $5026 \times 866$  design matrix **A** were given non-zero values ..."

Today the size of the matrices should be no problem on a standard PC.

On SVD:

"... indeed the earliest text-book we found referring to the technique was published in 1983."

G. H. Golub and W. Kahan: Calculating the singular values and pseudo-inverse of a matrix. SIAM J. Numer. Anal Ser. B 2 (1965) 205-224

G.H. Golub and C. Reinsch, Singular value decomposition and least squares solutions, Numer. Math. 14 (1970) 403-420

J. H. Wilkinson and C. Reinsch, Handbook for Automatic Computation, Vol. II. Springer (1971) (with ALGOL code)

C. L. Lawson and R. J. Hanson: Solving Least Squares problems. Prentice-Hall (1974)

In the older psychometric literature, such decompositions were called Eckart-Young decompositions, after C. Eckart and G. Young, The approximation of one matrix by another of lower rank. Psychometrika 1 (1936) 211-218.

### Alignment of the upgraded VDET at LEP2 (ALEPH)

A. Bonissent et al., Alignment of the upgraded VDET at LEP2, ALEPH 97-116

A global  $\chi^2$  involving all  $864 = 144 \times 6$  degrees of freedom is built, using single tracks and vertex constraints. Selected information is used from the outer tracking.

Precise faces measurement are used to reduce the degrees of freedom, while allowing for parametrized distortions.

The method is shown to provide accurate results, even with a limited number of events.



Local alignment of the VDET in the  $r\phi$  view. The dashed lines show the nominal position. The end wafers are drawn with solid lines, the other with dashed lines. Displacements are amplified by a factor of 100.

### Delphi: final alignment of the Barrel Silicon Tracker

P. Brückmann de Renstrom: The Final Alignment of the Barrel Silicon Tracker at LEP2, Delphi 2004-047 TRACK 098

Fitting the alignment parameters to the track-hit residual distributions (with reference layers).

Individual parametrizations of hit residuals fitted using MINUIT, including flat background from miss-associated hits.

Iterative sequence of applying geometry corrections.



A fit to the di-muon residual distribution.

$$\begin{split} \delta_z &= C + D \times \cot \theta + e \times r \times \cot \theta^2 \\ &\oplus \text{ flat background} \end{split}$$

Final average single-layer resolution in  $R\phi$  direction better than 8  $\mu$ m.

## Alignment using Kalman filter

CMS: About 20 000 silicon sensors, with resolution 10  $\mu$ m to 40  $\mu$ m. Expected precision from mechanical mounting and LASER beam alignment worse than intrincic resolution – alignment using tracks is necessary.

Extension of the Kalman filter method of track fitting, with updating the current alignment parameters after each track.

R. Frühwirth, T. Todorov and M. Winkler, Estimation of Alignment Parameters, using the Kalman Filter with annealing, CMS Note 2002-008

R. Frühwirth, T. Todorov and M. Winkler, Estimation of Alignment Parameters using the Kalman Filter with annealing, Journal of Physics G: Nuclear and Particle Physics 29 (2003) 561–574

Update formulae for global parameters  $\boldsymbol{a}$  and their covariance matrix  $\boldsymbol{E}$ :

$$oldsymbol{a}_1 := oldsymbol{a}_0 + oldsymbol{E}_0 oldsymbol{D}^{ ext{T}} oldsymbol{W} oldsymbol{[m-f(p_0, a_0]]} with \quad oldsymbol{W} = igg[oldsymbol{V} + oldsymbol{H}oldsymbol{C}_0 oldsymbol{H}^{ ext{T}} + oldsymbol{D}oldsymbol{E}_0 oldsymbol{D}^{ ext{T}} igg]^{-1} \ oldsymbol{E}_1 := oldsymbol{E}_0 - oldsymbol{E}_0 oldsymbol{D}^{ ext{T}} oldsymbol{W} oldsymbol{D} oldsymbol{E}_0$$

... requires some matrix operations.

Convergence to local minima not excluded: introduction of "annealing" = gradually stepping up the weights of the observations in the course of the estimation process (not "simulated annealing").

Simulated annealing is a method, where, depending on a temperature T and some BOLTZMANN factor, a step can result in an *increase* of an energy function.

Danny Hindson, A Robust Procedure for Alignment of the ATLAS Inner Detector Using Tracks, Thesis, Oxford (2004)

R. Harr et al., Tracking Detector Alignment Using Constrained Vertex Fits, IEEE Transactions on Nuclear Science, Vol 41 (1994) 796

K. Abe et al., Design and performance of the SLD detector, a 307 Mpixel tracking system, Nucl. Instr. and Methods A 400 (1997) 287–343

R. McNulty, T. Shears and A. Skiba, A Procedure for the Software Alignment of the CDF Silicon System, CDF/DOC/TRACKING/GROUP/5700

B. Mours et al., The design, construction and performance of the ALEPH silicon vertex detector. Nucl. Instr. and Methods A 379 (1996) 101–115

V. Karimäki, A. Heikkinen, T. Lampén and T. Lindén, Sensor alignment by tracks, CMS Conference Report, CHEP03, La Jolla, California, March 24-28, 2003

Different types of alignment algorithms are in use in HEP experiments:

• Simple residual methods (which are simple, but may require many iterations).

"The handle on alignment comes from the **residual distributions**." [Atlas report]

- Sequential and iterative ("classical") methods: hierarchy of parametrizations of residual distributions (residuals between the track extrapolation and the recorded cluster) and overall fit, iteratively.
- Simultaneous fits to (global) alignment parameters and (local) track parameters (MILLEPEDE).

Some recommendations:

- integration of alignment into reconstruction code;
- simultaneous use of all relevant detectors, but not too many parameters (orthogonality!);
- simultaneous use of several (all) types of events and data physics data: single tracks, vertices, invariant-mass constraints, and cosmics, overlaps ...

### Summary

Present and future detectors are very large and very accurate –

the requirements for good alignment and calibration are increasing.



Sometimes an observer has the impression that the alignment of certain objects can be improved with appropriate methods.

Data, computing power and also methods should be available to reach the goal.



Millepede: any of numerous herbivorous nonpoisonous arthropods having a cylindrical body of 20 to 100 or more segments most with two pairs of legs.

# Appendix. Matrix inversion

Matrix inversion is an essential part of minimisation. Inversion is a  $n^3$  process and can be time consuming for large matrix dimension. Complete matrix inversion fails for *singular* matrices and is *inaccurate* for almost singular matrices.

From HEP publications:

"For experiments with many data points, the inversion of such large matrices may lead to numerical instabilities, in addition to being time-consuming."

"Minimizing  $\chi^2$  ... is impractical because it involves the inversion of the measurement covariance matrix which, in global fits, tends to be very large."

Matrices to be inverted in statistical computation are *symmetric* and represent covariance/Hessian matrices:

 $\Delta \boldsymbol{a} = \boldsymbol{C}^{-1} \boldsymbol{b}$   $\boldsymbol{C}^{-1} =$  covariance matrix of the parameters

#### The result will reflect the statistical properties of data and model!

- The storage and the computation can make use of the symmetry ( $\rightarrow$  factor 1/2).
- Matrices with highly correlated parameters are almost singular.
- Highly correlated parameters should be avoided (and are not meaningful). A strategy to get relevant results in difficult cases can be used.

Time for inversion of a *n*-by-*n* matrix is expected to be  $\approx \mathcal{O}(n^3)$ .

n =	RINV	SMINV	HHLROT	unit
10	13.5	7.4	29.4	$\mu sec$
100	7.7	2.6	11.6	msec
1000	12.4	3.3	16.4	sec
3162	17.9	2.4	28.7	$\min$
words	$n^2$	$1/2 \ n^2$	$3/2 \ n^2$	

Inversion with n = 25000 will take  $\approx$  one day (SMINV) and requires 1.25 GB. Atlas benchmark on 16-proc cluster: 95 hours for n = 8000

Atlas benchmark on 10-proc cluster: 95 nours for n = 8000in double precision.



#### 2.6 MHz Pentium with 512 MB; single precision computation.

Alignment for tracking detectors

Check of accuracy based on  $\mathbf{V}' = (\mathbf{V}^{-1})^{-1}$ . Plots show  $\log_{10}$  of difference versus sum of elements; lines correspond to  $10^{-2}$   $10^{-4}$   $10^{-6}$ .



Highest precision by RINV (Cern); lowest precision for inversion by diagonalisation.

### Matrix programs

- **RINV** Cern-Library program for full matrices, using triangular factorization with row interchange (special code for  $n \leq 3$ ). Returns flag for singularity, but singularity will often go undetected.
- **SMINVX** Special Gauss-Jordan algorithm for **symmetric matrices** with pivot selection on diagonal, and singularity detection.
- **SMINV** Same as **SMINVX**, but with index calculation avoiding integer multiply and up to a factor of 3 faster.

Algorithm: use always largest pivot element (on diagonal), but avoid elements with very large  $(V)_{jj} \cdot (V^{-1})_{jj}$ . Stop inversion if no acceptable pivot can be found and clear corresponding matrix elements, i.e. invert the largest possible submatrix, and return zero correction for remaining parameters.

**HHLROT** Diagonalization (eigenvalues + eigenvectors) by Householder transformation followed by diagonalization of tridiagonal matrices. Allows to recognize insignificant components of the solution from eigenvalues.

The global correlation coefficient,  $\rho_j$  is a measure of the total amount of correlation between the *j*-th parameter and *all* the other variables. It is the largest correlation between the *j*-th parameter and every possible linear combination of all the other variables.

$$\rho_j = \sqrt{1 - \frac{1}{(\mathbf{V})_{jj} \cdot (\mathbf{V}^{-1})_{jj}}} \quad \text{and} \quad (\mathbf{V})_{jj} \cdot (\mathbf{V}^{-1})_{jj} = \frac{1}{1 - \rho_j^2}$$

Precision of result for matrix, made singular with rank defect of 1.



RINV (Cern) fails without result. Other algorithms have still useful result for 999 by 999 submatrix, with unchanged precision.

What is the best method to solve a very large least squares problem?

#### Orthogonal decomposition by singular value decomposition (SVD)

Advantage: with single-precision arithmetic as accurate as other methods with double-precision. Disadvantage: requires the full input matrix in memory.

Another method which allows "sequential" input of data:

Transformation by Householder triangularization to triangular form without requiring that the entire matrix be in computer storage.

C. L. Lawson and R. J. Hanson: Solving Least Squares problems. Prentice-Hall (1974)

## Contents

1.	Alignment and calibration Alignment/calibration
2.	Alignment of a toy detector
	First attempt based on residuals
	Alignment using iterative track fitting
	Result from the first attempt
	First attempt – Discussion
	Results from the second attempt
	Use of higher mathematics?
	Results from a simultaneous fit
	Determination of drift velocities
3.	Linear least squares and matrix equations

The standard	linear	least	squares	method	Ι		•
The standard	linear	least	squares	method	Π	•	•

<b>4</b> .	The MILLEPEDE algorithm	16
	Inversion by Partitioning	17
	Partitioning the solution vector	18
	The MILLEPEDE formalism	19
	Constraints in MILLEPEDE	20
	Millepede	21
	The MILLEPEDE program	22
5.	Drift chamber calibration	23
	H1 Determination of 1400 parameters	24
	H1 Drift chamber alignment and calibration	25

6. Vertex detector alignment	26
H1 calibration	27
Internal alignment of the SLD vertex detector	28
Mathematical aspects	31
Alignment of the upgraded VDET at LEP2 (ALEPH)	32
Delphi: final alignment of the Barrel Silicon Tracker	33
Alignment using Kalman filter	34
Further references	35
7. Conclusions and summary	36
Summary	37
	38
Appendix. Matrix inversion	39
Matrix inversion – timing	40
Matrix inversion – accuracy	41
Matrix programs	42
Singular matrix and inversion	43
Solution of very large least squares problems	44
Table of contents	<b>45</b>